

DECLARATION OF CONFORMITY

IN ACCORDANCE TO ISO/IEC GUIDE 22
FOR A

LIGHT SOURCE

MANUFACTURER: Rosemount Analytical, Inc.
1201 North Main Street
Orville, OH 44667
Phone: 330-684-4407
Fax: 330-684-4431

MODEL #: AOTF-NIR Light Source

REPORT #: 2251-01

DIRECTIVE: Low Voltage Directive (73/23/EEC)

STANDARD: Safety requirements for electrical equipment for measurement, control and laboratory use – Part 1: General requirements, EN 61010-1: 1993 edition

TEST FACILITY: F-Squared Laboratories
9890 Main Street
Damascus, MD 20872
Phone: 301-253-4500
Fax: 301-253-5179

The AOTF-NIR Light Source is in effective conformance to the Directive and Standard referenced above.

Authorized by:



Date: 4-12-00

Name:

IAN MACLEOD

Title: V.P.