

EC DECLARATION OF CONFORMITY

Document number: RAE/BINOS100 2M-E8

Date: February 2008

We,

Emerson Process Management GmbH & Co. OHG

located at

Industriestrasse 1, D-63594 Hasselroth, Germany

declare under our sole responsibility that our gas analyzer, type

BINOS 100 2M

to which this declaration relates is in conformity with the provisions of:

89/336/EEC EMC Directive (changed by directive 91/263/EEC 92/31/EEC and 93/68/EEC)
with the application of the harmonized standards:

EN 61326-1:1997 Electrical equipment for measurement, control and laboratory use -
+ A1:1998 + A2:2001 EMC requirements
+ A3:2003

73/23/EEC Low Voltage Directive (changed by directive 93/68/EEC)
with the application of the harmonized standards:


EN 61010-1:2001 Safety requirements for electrical equipment for measurement, control and
laboratory use
Part 1: General requirements

97/23/EC Pressure Equipment Directive

This analyzer has been designed and manufactured considering article 3,
paragraph 3 of the above mentioned directive and therefore CE marking
does not refer to this directive.

BINOS 100 2M does include any ½ 19" BINOS 100 2M analyzer (both rack-mountable or
table-top version).

Hasselroth, February 2008


.....
(Signature)

Andy Kemish
(Name)

VP Rosemount Analytical Europe
(Function name)

ROSEMOUNT[®]
Analytical


EMERSON[®]
Process Management

This declaration confirms the compliance with announced directives but does not include the assurance of properties.
The safety and installation instructions of the documentation have to be followed.