

Certificate of Conformance

- ▶ **Product Type:** Gas Analyzer
- ▶ **Model Designation:** Binos 100 2M
- ▶ **Manufacturer:** Fischer Rosemount GmbH & Co.
Industriestraße 1
D - 63594 Hasselroth
- ▶ **Certificate Number:** C98WI03

A representative sample of the unit identified above was evaluated, tested, and found to be in compliance with the following requirements:

IEC 1010-1:1990+A1:1992
EN 61010-1:1993+A2:1995

- ▶ **Evaluation Unit:** NONE
- ▶ **Testing Laboratory:** Intertest Systems GmbH
Berliner Straße 207
D - 65205 Wiesbaden
- ▶ **Report Number:** 98WIS053
- ▶ **Authorized Certification Mark:** IS - Mark
- ▶ **Factory Surveillance:** Yes


Dipl.-Ing. Wladislaw Kreuzschner
Reviewing Engineer


Dipl.-Ing. Holger Schmieid
Reviewing Engineer

- ▶ **Issued:** May, 26th 1998
- ▶ **Place:** Wiesbaden